

CE technical document	Document Number	Q/DL-JS-CE-003-01
	Version / Revision	01
Medical Suction Unit	Page	1 of 1

EC Declaration of Conformity

Manufacturer:

Jiangsu Dynamic Medical Technology Co., Ltd
 No.108 Xingpu Road, Lujia Town, Kunshan,
 City, Jiangsu Province, 215331.
 PEOPLE'S REPUBLIC OF CHINA

whose single Authorized Representative:

Shanghai International Holding Corp Gmbh
 (Europe)
 Eiffestrasse 80, 20537 Hamburg, Germany

We, the manufacturer, herewith declare that the products

Medical Suction Unit

Model: DS series.

DS1000、DS2501M、DS3701M、DS3701CS、DS3701CS-1、DS3701CS-2011、DS7501CS、
 DS01、DS02、DS03、DS04、DS06、DS08、DS10、DS301、DS302、DS502、DS504、
 DS506、DS508、DS510、DAS001、DAS002、DS803、DS805、DS808、DS810、DS910、
 DS915、DS920、DS930、DS940、DS950

UMDNS-Code: 10-212

meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class I (active) according to Rule 12 of Annex IX of the Directive 93/42/EEC. It bears the mark



The product concerned has been manufactured according to Annex V.3 of Directive 93/42/EEC.

following the procedure relating to the Annex V.3 of Directive 93/42/EEC, and complying with EN standards of:

EN 980:2008, EN 1041: 2008, EN ISO 14971:2009, EN 60601-1:2005, EN 60601-1-2:2015, EN 1640:2009, EN 15223-2012, ISO11143:2008.

The above mentioned declaration of conformity is exclusively under the responsibility of

Jiangsu Dynamic Medical Technology Co., Ltd

Jiangsu, China, Mar 22, 2020
Place, date

 General Manager
Legally binding signature, Function

